
5. Immediately touch the tip of the test strip to the drop of blood. The blood should be pulled into the test strip through the tip. Do not put blood on top of the test strip.



6. The meter counts down from 5 to 1 and then your test result is displayed and automatically stored in the meter memory.



Note:

1.The meter and lancing device are for single patient use. Do not share them with anyone including other family members! Do not use on multiple patients.

2.All parts of the kit are considered biohazardous and can potentially transmit infectious diseases, even after you have performed cleaning and disinfection.

3.You should reference the user manual for cleaning and disinfection procedures.

Manufactured for:
Smart Meter, LLC
6206 Benjamin Rd.
Ste. 314 Tampa, FL 33634

Contact Smart Meter Customer Support
Monday-Friday 9:00AM-5:00PM EST 844-445-8267
Please contact your healthcare provider for assistance after normal business hours

Number:
Date:

Troubleshooting Guide

Refer to your User Manual for complete troubleshooting and error information, or call Customer Support: 844-445-8267 Monday-Friday 9:00AM-5:00PM EST.
Please contact your healthcare provider for assistance after normal business hours.

	What It Means	What You Should Do
E 1	Blood or control solution was applied to the test strip before the flashing drop appeared on the display.	Discard the test strip and repeat the test with a new test strip. Wait until you see the flashing blood drop on the display before testing.
E 2	The meter is sensing a used or contaminated test strip.	Discard the test strip and repeat the test with a new test strip. Wait until you see the flashing blood drop on the display before testing.
E 3	Incorrect test strip.	Discard the test strip and repeat the test with a new test strip. Make sure that you are using a iGlucose®Essential test strip.
E 4	Incorrect sample.	Discard the test strip and repeat the test with a new test strip. Make sure that only human capillary blood and iGlucose®Essential control solution are used for the test.
E 5	Temperature out of range.	Move to an area that is within the operating range for the meter. Let the meter adjust to this temperature for 20 minutes before performing a test.
E 6 E 7	Potential hardware error.	Restart the meter. If the problem continues, contact Customer Support
E 8	A test strip inserted while the meter was connected to a computer or wall outlet.	When the charge is completed (about 3 hours for charging an empty battery), remove the Micro USB cable from the meter, and then take a test.
E 10	Insufficient sample.	Repeat the test and apply enough sample to fill the test strip check window.
E 11	Running out of battery.	Charge the battery.
	Data not transmitted	Please check network connection. Then discard the test strip and repeat the test with a new test strip, if the problem continues, contact Customer Support: 844-445-8267 Monday-Friday 9:00AM-5:00PM EST. Please contact your healthcare provider for assistance after normal business hours.
	No signal	Restart the meter. Please check if your SIM data is used up. If the problem continues, contact Customer Support.
HI	Test result is above 600 mg/dL.	Wash and dry your hands and the test site well.Repeat the test using a new test strip. If your result still flashes HI, contact your health care professional as soon as possible.
LO	Test result is below 20 mg/dL.	Wash and dry your hands and the test site well.Repeat the test using a new test strip. If your result still flashes LO, contact your health care professional as soon as possible.

iGlucose®Essential

Made in China

Blood Glucose Meter


QUICK START GUIDE

MODEL: SMBGM-TMB-2282-G
REF: TeleBGM 2282-G


Read your User Manual for complete instructions before performing a blood glucose test.
For Single-patient use only. Do not share.

SMARTMETER
Engaging Health Data

Your Meter and Test Strip



Your Lancing Device




Prepare the Lancing Device


Important:

Prior to testing, wipe the test site with an alcohol swab or soapy water. Use warm water to increase blood flow if necessary. Then dry your hands and test site thoroughly.


1. Unscrew the lancing device cover from the body of the lancing device. Insert a sterile lancet into the lancing device.




2. Push the lancet firmly in the lancing device and twist the safety tab of the lancet until it loosens, then pull the safety tab off the lancet.




3. Carefully screw the cover back onto the lancing device.



4. Adjust the puncture depth by rotating the lancing device cover.



5. Pull the cocking barrel back to set the lancing device.



Performing a Blood Glucose Test

1. Remove a test strip from the strip vial. Tightly close the vial cap immediately after you have removed the test strip.


2. Insert the test strip into the meter in the direction of the arrows. The meter will turn on automatically.



3. A symbol of a flashing blood drop will appear letting you know the meter is ready to test.



4. Press the lancing device against the side of your finger with the cover resting on the finger. Push the release button to prick your fingertip.



颜色：CMYK

Description	TeleRPM Blood Glucose Meter QRG(En)	Code No.	
Material	128g铜版纸+水性上光，Z字折，四色	Size	420mm×105mm
Deigned by		Checked by	
Approved by		Effective Date	